



SUNRISE MEDICAL CARES ABOUT YOUR SAFETY. Please read your owner's manual for safe operating instructions and maintenance guidelines.

CALIBRATING YOUR M90

The wheels should be calibrated during the initial installation or if installing the wheels on a different chair.

The calibration process sets the M90's slope and hill thresholds based on the wheelchair, optimizing performance for the rider.

CALIBRATION INSTRUCTIONS

IMPORTANT: Do not move the chair during calibration.

- Engage the wheelchair wheellocks, and remove the battery on the M90 wheel.
- Press and hold either button on the wheel while re-inserting the battery.
- 3. When the lights turn ORANGE release the button.
- 4. Wait until the lights turn WHITE then GREEN.
- When the lights are GREEN, remove and re-insert the battery. The wheel will beep.
- 6. Repeat steps 1 5 on the other wheel.
- Remove and swap wheels to the opposite sides and repeat steps 1 – 6.

NOTE Re-calibrating the wheels is the first step for any troubleshooting with the M90.



POWFRING ON/OFF

When the batteries are installed, the wheels are powered on. The wheels will automatically search for the controller via a BTE connection.



NOTE Buttons on the M90 wheels are used as indicators and trouble shooting, not as an On/Off.

THE DUAL CONTROLLER

Your M90 comes with two driving options:





CONNECTING THE CONTROLLER

Press the Controller Power button once.

The Power Indicator LED (2) will flash white while connecting. (This may take up to 5 seconds.)

When connected, you will hear a beep, and the Power LED (2) will turn solid green.

With the wheels engaged, hold the toggle forward to enter drive mode. Both wheels will lock into place with a click, and the status light (3) will turn green when ready.



2- CONTROLLER POWER INDICATOR

1 - WHEEL BATTERY INDICATOR

90 - 100%

60 - 90%

30 - 60%

0 15 - 30%

5 - 15%

<5%

- CONNECTED, LOCKED Flashing = Awaiting Connection
- CONTROLLER ACTIVE Charge = >15%
- CHARGE <15% Flashing = Charge <5%</p>
- LOW CHARGE
 Unable to be used
- CHARGED
 Flashing = Charging

3- STATUS INDICATOR

- WHEEL HUBS NOT ENGAGED
- WHEEL HUBS ENGAGED Flashing = Cruise Control engaged
- Thermal Protection 1 –
 Outdoor Mode Lock-Out
- Thermal Protection 2 –
 Controller Locked. Active braking
 for 5 seconds
- Thermal Protection 3 –
 System Lock-Out
- CHARGED
 Flashing = Charging

4- PROFILE / DRIVE MODE INDICATOR

- JOYSTICK DRIVE MODE
 Solid = Indoor Drive Profile
 Flashing = Outdoor Drive Profile
- SPEED CONTROLLER WITH HANDRIM STEERING DRIVE MODE

Solid = Indoor Drive Profile Flashing = Outdoor Drive Profile

FUNCTIONS	JOYSTICK CONTROLLER	SPEED CONTROLLER with HANDRIM STEERING
START / STOP	Push Joystick toggle forward to start. Let go of the toggle to stop. Handrims do not affect speed.	Tap Speed Controller toggle to start. Push speed controller toggle down or use handrims to stop gradually.
SPEED CONTROL	Push Joystick toggle forward or backward to proportionately accelerate or slow down.	Tap Speed Contoller toggle to increase or decrease speed step by step, up to level 5 in both profiles.
STEERING	Move Joystick toggle in desired direction to steer.	Steer manually using the handrims.
CRUISE CONTROL	Press the Joystick toggle forward until desired speed is reached, then press down to activate. Steer by moving the Joystick toggle right or left. Exit by moving the Joystick toggle forward or back, or pressing down.	Tap Speed Contoller toggle forward or backward to increase or decrease speed level. Steer manually with handrims. Exit by pressing Speed Contoller toggle down or by grabbing and holding the handrims.
MANUAL ASSIST	N/A	A gentle boost that engages after a stop and allows you to reposition your chair using the handrims.
DRIVE PROFILE Indoor/Outdoor	Select using the M90 App or the Dual Controller	Select using the M90 App or the Dual Controller
DRIVE SETTING	Choose speed range (Base / Standard / Max) using the M90 App.	Choose speed range (Base / Standard / Max) using the M90 App.

SAFE DRIVING TIPS

Start Slow. We recommend anti-tips when operating the M90.

Start in Indoor mode in a safe, open environment. Practice stopping at specific points, turning while stationary, and turning while in motion.

HILLS

Begin on a moderate hill with a gradual slope and a safe runout, avoiding immediate turns at the bottom. The M90 provides braking and descent control.

Adjust your body position to maintain balance and prevent tipping – leaning forward to go up or backward to descend. Ensure your trunk is properly supported, if necessary.



SCAN HERE FOR SET-UP AND USE VIDEOS

Or, scan to view this document in other languages.

Escanea para ver este documento en ESPAÑOL.

Scannez pour afficher ce document en FRANÇAIS.

Scannen, um dieses Dokument auf **DEUTSCH** angezeigt zu bekommen.

Inquadra qui per vedere il documento in ITALIANO.

Scan om dit document in het ${f NEDERLANDS}$ te bekijken.



www.SunriseMedical-Group.com



