

## Bill of Materials


<b>Company:</b>	Sunrise Medical S.L. Polígono Bakiola, 41 48498 Arrankudiaga Vizcaya / ESPAÑA
-----------------	----------------------------------------------------------------------------------------

<b>Product:</b>	Q100R & Q200R
-----------------	---------------

The following materials (that could contact the user) are used in the above products:

DESCRIPTION	MATERIAL
Seat Sling	Polyester Oxford Weave, 300D
Seat Frame	Powder coated Steel
Std Seat Cushion - Cover	Polyester Oxford Weave, 300D
Std Seat Cushion - Foam	Polyester Foam, 50m thick, Density 50kg/m3 open cell
Backrest Upholstery	Polyester Oxford Weave, 300D
Backrest Frame	Powder coated Steel
Armrest Pad	Polyurethane
Armrest Receiver	Powder coated Steel
Legrest	Powder coated Steel
Footplate	Powder coated Aluminium
Hanger Release lever & Cover	ABS, ISO1043
Shrouds	ABS, ANC160 UL94V0
Joystick	Santoprene 101-80
Motor Brake Lever	NBR

Mr. D.S.Davies – Director Project Management and R&D Europe	A	29.10.19
<b>Approval Name and Function</b>	<b>Revision</b>	<b>Approval Date</b>


<b>Signature</b> (Sunrise Medical Approval representative)

GMS Form Number:	Revision: <b>B</b>	Effective Date: <b>01.02.2010</b>
Form Owner: Heads of Engineering	Form Approver: Global Head of Engineering	GMS Change Number:
Page 1 of 1		