



NO.212400487495200



运输危险性鉴定书

Hazard Classification and
Identification Report for Transport of Goods

危险品

样品名称： 电动轮椅 Q50R (内置可充电锂电池组18650 7S12P 24V 30Ah 720Wh)

Sample name: Power Wheelchair Q50R (Containing Rechargeable Lithium Battery Pack 18650 7S12P 24V 30Ah 720Wh)

委托单位： 江苏英科医疗制品有限公司
JIANGSU INTCO MEDICAL PRODUCTS CO., LTD.

生产单位： 江苏英科医疗制品有限公司
JIANGSU INTCO MEDICAL PRODUCTS CO., LTD.



Witness Better Life

SICIT

上海化工院检测有限公司

SHANGHAI INSTITUTE OF CHEMICAL INDUSTRY TESTING CO., LTD.



运输危险性鉴定书

Hazard Classification and Identification Report
for Transport of Goods

NO. 212400487495200

Page 1/3

样品名称 Sample Name	中文 Chinese	电动轮椅 Q50R (内置可充电锂电池组18650 7S12P 24V 30Ah 720Wh)		
	英文 English	Power Wheelchair Q50R (Containing Rechargeable Lithium Battery Pack 18650 7S12P 24V 30Ah 720Wh)		
委托单位 Applicant		江苏英科医疗制品有限公司 JIANGSU INTCO MEDICAL PRODUCTS CO.,LTD.		
生产单位 Manufacturer		江苏英科医疗制品有限公司 JIANGSU INTCO MEDICAL PRODUCTS CO.,LTD.		
检验方法、程序 Inspection method and procedure		国际海事组织《国际海运危险货物规则》(2022版) IMO International Maritime Dangerous Goods Code (2022 Edition)		
样品外观 Sample appearance		黑色塑胶及金属外壳 Black Plastics cement and metal shell		
包装件信息 Package information		/		
序号 NO.	电池种类 Battery type	型号 Model	容量Capacity /锂含量Li content	放置方式 Placement
1	可充电锂离子电池组 Rechargeable Li-ion battery	18650 7S12P	30Ah 720Wh	安装在车辆内 Contained in vehicle
鉴定结论 IDENTIFICATION CONCLUSION	1. 危险性识别(Hazards identification) 杂项。 Miscellaneous.			
	2. 海运按照国际海事组织《国际海运危险货物规则》办理的类项(Suggestion according to IMO IMDG Code) Shipping Name:Battery-powered vehicle Class or Division:9 UN Number:UN3171			
	3. 包装要求(Packaging requirements) 按UN3171要求办理。 Suggestion according to UN3171.			
	检验日期: 2024-01-05 签发日期: 2024-01-05 生效日期: 2024-01-05 Inspection Date: Issue Date: Effective Date:			
备注 Comment		/		


批准
Approver: 王宗审核
Checker: 董学胜主检
Appraiser: 孙清

运输危险性鉴定书

Hazard Classification and Identification Report
for Transport of Goods

NO. 212400487495200

Page 2/3

序号 No.	检验结果及其他事项 Inspection results and other things
1	<p>本鉴定书所述锂电池按照《国际海运危险货物规则》(2022版) 2.9.4.5规定的质量管理体系进行制造。 Lithium cells and batteries listed in this report were manufactured under the quality management program described in IMDG CODE 2022 EDITION 2.9.4.5.</p>
2	<p>本鉴定书所述锂电池已通过联合国《试验和标准手册》第III部分38.3小节相应测试要求。 Lithium cells and batteries listed in this report are of the types proved to meet the requirements of each applicable test in the UN Manual of Tests and Criteria, Part III, sub-section 38.3. UN38.3试验概要编号 The UN38.3 Test Summary No. (s) 812100200754310 详细信息请扫描右侧二维码。 Please scan the QR code on the right for more information.</p> 
3	<p>电池应安装并固定在车辆的电池仓内，并有适当的防止损坏及短路的保护措施。 The batteries should be installed and securely fastened in the battery holder of the vehicle and be protected in a suitable manner to prevent damage and short circuits.</p>
4	<p>本规则中有关标记、标签、标牌和海洋污染物的规定不适用。 The provisions of this Code relevant to marking, labelling, placarding and marine pollutants shall not apply.</p>
5	/
6	/
7	/
-验证码:966082-	

检验
专用

运输危险性鉴定书

Hazard Classification and Identification Report
for Transport of Goods

NO. 212400487495200

Page 3/3



Li-ion Battery Park

Product model: 18650 7S12P

Normal voltage: 24V

Normal capacity: 30Ah 720Wh 71INR19/66-12

Charging voltage: 29.4V

Jiangsu Intco Medical Products Co., LTD



BBS243020210107021



ROHS

IEC 62133

UN38.3

(14)

报告结束